



FINAL REPORT

MEM ELUTION

PROCEDURE NO. STP0032 REV 03

LABORATORY NO. 439112.1 AMENDED

PREPARED FOR:

KEITH MOELLER  
AMERICAN BIOTECH LABS  
80 W. CANYON CREST RD.  
ALPINE UT 84004

SUBMITTED BY:

NELSON LABORATORIES, INC.  
6280 S. REDWOOD RD.  
SALT LAKE CITY UT 84123-6600  
801-290-7500

MEM ELUTION

LABORATORY NUMBER:	439112.1 Amended
PROCEDURE NUMBER:	STP0032 REV 03
SAMPLE SOURCE:	American Biotech Labs
TYPE OF TEST:	Solid, Gel
SAMPLE IDENTIFICATION:	Refer to Table 1
DEVIATIONS:	None
CELL LINE:	Mouse Heteroploid Connective Tissue (L-929)
INCUBATION PERIOD:	72 ± 3 hours at 37 ± 1°C
METHOD OF SCORING:	Cytopathic Effect (0-4)
AMOUNT TESTED/SAMPLE EXTRACT:	3 g / 15 mL
SAMPLE RECEIVED DATE:	20 Aug 2008
LAB PHASE START DATE:	21 Aug 2008
LAB PHASE COMPLETION DATE:	26 Aug 2008
REPORT ISSUE DATE:	27 Aug 2008
STUDY COMPLETION DATE:	27 Aug 2008
AMENDED REPORT ISSUE DATE:	03 Nov 2008

AMENDMENT JUSTIFICATION:

At the request of the sponsor, the final report was changed from short format to long format.

INTRODUCTION:

The MEM Elution test was designed to determine the cytotoxicity of extractable substances. An extract of the sample was added to cell monolayers and incubated. The cell monolayers were examined and scored based on the degree of cellular destruction.

ACCEPTANCE CRITERIA:

The United States Pharmacopeia & National Formulary states that the sample meets the requirements if the reactivity grade is not greater than grade 2 or a mild reactivity. The AAMI/ISO 10993-5 standard states that the overall assessment of the results shall be made by capable persons based upon the data and results. Nelson Laboratories acceptance criteria was based upon the negative and media controls receiving "0" reactivity grades and positive results receiving a 3-4 reactivity grades (moderate to severe).

**PROCEDURE:**

The amount of test material extracted was based on ANSI/AAMI/ISO and USP surface area recommendations or weight (0.20 g/mL extract fluid for polymers and plastic). The sample was extracted for 24-25 hours at  $37 \pm 1^\circ\text{C}$  in 1X Minimal Essential Media with 5% calf serum. Positive (Latex Natural Rubber) and negative (Polypropylene Pellets) controls were extracted and included in the assay. A blank of extraction media (media control) was also included in the assay.

Multiple well cell culture plates were seeded with a verified quantity of L-929 cells and incubated until 80-90% confluent. The cell culture media was removed from the plates. The test extracts were filtered and the appropriate amount of extract was added to each well on the cell culture plates. Each extract was tested on three wells of cells. The cells were incubated at  $37 \pm 1^\circ\text{C}$  with  $5 \pm 1\%$   $\text{CO}_2$  for  $72 \pm 3$  hours.

The cell monolayers were examined microscopically. The wells were scored as to the degree of discernable morphological cytotoxicity on a relative scale of 0 to 4:

CONDITIONS OF ALL CULTURES	REACTIVITY	GRADE
No cell lysis, intracytoplasmic granules.	NONE	0
Not more than 20% rounding, occasional lysed cells.	SLIGHT	1
Not more than 50% rounding, no extensive cell lysis.	MILD	2
Not more than 70% rounding and lysed cells.	MODERATE	3
Nearly complete cell destruction.	SEVERE	4

The results from the three wells were averaged to give a final cytotoxicity score.

**RESULTS:**

The results are summarized in Table 1. The test is acceptable if all three of the negative control and medium control test wells have a score of 0 and all three of the positive control test wells have a score of 3 or higher.

The sample meets USP requirements if none of the cell culture exposed to the sample shows greater than a mild reactivity (grade 2).



American Biotech Labs  
Lab Number 439112.1 Amended

MEM Elution

STATEMENT OF UNCERTAINTY:

If applicable, the statement of uncertainty is available to sponsors upon request.

  
\_\_\_\_\_  
Technical Reviewer

 for  
\_\_\_\_\_  
Bobbi Rushton-Castro  
Study Director

  
\_\_\_\_\_  
Amended Report Date

jzw

TABLE 1. Results

IDENTIFICATION	SCORE #1	SCORE #2	SCORE #3	AVERAGE
Negative Control	0	0	0	0
Media Control	0	0	0	0
Positive Control	4	4	4	4
429606 Lot #060908	1	1	1	1



American Biotech Labs  
Lab Number 439112.1 Amended

MEM Elution

All reports and letters issued by Nelson Laboratories, Inc. are for the exclusive use of the sponsor to whom they are addressed. These results relate only to the samples tested. Reports may not be reproduced except in their entirety. No quotations from reports or use of the corporate name is permitted except as expressly authorized by Nelson Laboratories, Inc. in writing. The significance of any data is subject to the adequacy and representative character of the samples tendered for testing. Nelson Laboratories, Inc. warrants that all tests are performed in accordance with established laboratory procedures and standards. Nelson Laboratories, Inc. makes no other warranties of any kind, express or implied. Nelson Laboratories, Inc. expressly states that it makes no representation or warranty regarding the adequacy of the samples tendered for testing for any specific use of application, that determination being the sole responsibility of the sponsor. Nelson Laboratories' liability for any loss or damage resulting from its actions or failure to act shall not exceed the cost of tests performed, and it shall not be liable for any incidental or consequential damages.